

# A Two-Stage Color Correction Algorithm for Facial Image

---

## ABSTRACT

**Aims:** Facial image analysis is the main direction of facial diagnosis objectification, where realistic colors in medical facial images are important clues for disease diagnosis. However, under different light sources and display devices, the same person's facial color information will have different display information may lead to different diagnostic results.

**Methodology:** Firstly, we conducted group experiments on the selection of color patches and selected 12 face-related color patches for the task of color correction of facial images. Next, to adapt to this task in different lighting environments, we utilize an adaptive white balance algorithm to adjust brightness as the first step. We selected the D65 illuminate environment as the standard lighting condition. Finally, we employed a polynomial regression algorithm based on ridge regression for color correction.

**Results:** The experimental results demonstrate that the average values of the chromatic distance of our method in all five light sources are less than 3.0. We also compare the ablation experiments of adjusting brightness.

**Conclusion:** In conclusion, our experimental results indicate that the proposed method has good consistency in color correction of facial images under different lighting conditions which is more suitable for clinical medical diagnosis.

*Keywords: Color Correction, Facial Image, White Balance, Regression model*

## 1 Introduction

With the continuous development of artificial intelligence technology, researchers have increasingly used computer-aided diagnosis in the field of clinical diagnosis. In traditional Chinese medicine (TCM), facial color inspection is an important diagnostic method, which conveys abundant valuable information about the physiological conditions and pathological changes of patients(1). However, even for the same face, the captured facial images taken by different cameras under different lighting conditions display different color appearances on the monitor(2). It may lead to different diagnosis results by clinical doctors, and thus harm the health of patients. Therefore, color correction of facial images is essential for computer-aided image analysis of clinical diagnosis.

Many scientific researchers apply digital image technology to realize the potential of

---

color correction in medical diagnosis. In the process of research, lots of methods have been investigated, such as polynomial-based methods (3–5), SVR-based methods (6–8), and neural network-based methods (2,9,10). The most common color correction method is to calculate the mapping relationship between the color space to be corrected and the target color space. Finlayson et al (3) used an alternating least squares algorithm for nonuniform intensity color correction in the 24-patch Macbeth color checker. Graham et al (4) applied root-polynomial regression to enhance color correction performance on real and synthetic data. Zhang et al (5) employed a polynomial transform-based correction algorithm to portray the facial images and facilitate quantitative analysis before feature extraction and classification. Besides, Zhuo et al. (6) designed a K-PLSR method for color correction under different lighting conditions. Bo et al (7) introduced support vector regression (SVR) to reproduce the colors of the nonlinear imaging system.

On the other hand, many researchers have exploited a significant amount of image data and neural network-based correction methods to achieve the consistency and stability of captured images and extracted features. Zhou et al (9) employed an SA-GA-BP neural network-based color correction algorithm for TCM images. They selected several colors that are similar to the tongue and skin to reduce the complexity of training processing. Zhang et al. (10) proposed a MEC-BP-Adaboost neural network-based color correction algorithm for color images. Lu et al (2) designed a two-phase deep color correction network and synthesized about 2000 images as a training set to adapt to different work environments for color appearance. However, it is difficult to obtain a tremendous amount of facial images because of marking costs and patient privacy. If a large amount of training samples is lacking, this algorithm is not suitable for face image correction (11).

Images captured under different lighting conditions will have different visual appearances depending on the illumination and color temperature. According to the recommendations of the International Commission on Illumination (CIE), a D65 light source with a color temperature of 6500K can be selected as the standard lighting environment to simulate daylight (2). Therefore, some scholars carried out image color correction tasks for color deviation under different lighting conditions. Akazawa et al. (12) proposed a three-color balance adjustment for color constancy correction. Yan et al. (13) designed an adaptive white balance method motivated by their statistical prior to mitigate the impact of color deviation. Huang et al. (14) employed a camera white balance conversion method based on triangle affine transformation and used two different cameras to image the same Digital SG ColorChecker.

Motivated by the functional gaps in current facial color correction in the case of small samples, we proposed two-stage color correction for face color correction. Due to the influence of different lighting conditions, the color appearance of facial images varies over a wide range. The proposed method applies an adaptive white-balanced algorithm to adjust the brightness of the facial image. Then, it utilizes a polynomial-based method to learn the color mapping model between the distorted and target color images for color correction.

This study demonstrates three aspects as follows:

(1) Instead of using a single correction method, we combined a white-balanced algorithm and a polynomial-based method for color correction under different illuminate environments.

(2) We tested group experiments on the selection of color patches in 24 Xrite ColorChecker and selected 12 face-related color patches for color correction.

(3) Experiments conducted on five different lighting conditions, results indicate that the proposed method performs well for color correction tasks

The remainder of this paper is organized as follows. Details of the proposed method are described in Section 2. The experimental results are analyzed in Section 3The discussion is presented in Section 4 Finally, conclusions and future work are summarized in Section 5.

## 2 METHODOLOGY

### 2.1 Overview

This section demonstrates the proposed correction algorithm for medical facial images, which is illustrated in Figure 1. As it is shown, the input is a distorted image containing face and color patches. Firstly, the distorted image is adjusted overall brightness by the proposed adaptive white-balance-based algorithm. Next, the color space values are extracted from the processed image, and the face image is identified using 68-point face detection (15) and cropped. Then, the color correction algorithm is used to optimize the mapping relationship between the distorted color space values and target D65 color space values. Finally, the face image is input into the optimized color correction model to obtain the corrected image.




Figure 1. The pipeline of the proposed correction algorithm

The dataset of the images in this study was 24-bit images of face, including some background information, with a spatial resolution of 3840×2880 pixels, was collected from the digital laboratory of four diagnosis of TCM, Shanghai University of Traditional Chinese Medicine, Shanghai, China. The volunteers are selected from the students of Shanghai University of Traditional Chinese Medicine. Informed consent to publish identifying images has been obtained. The facial images were captured by specialized equipment in five different lighting conditions. The equipment includes two assisted light sources and a portable camera with photographing medium (see Figure 2). The photographing takes distance 23cm-26cm between the camera and face. These lighting conditions include D50 illumination (1337lx, 5186k), D65 illumination (1830lx, 6991k), D65 illumination with warm light (5023lx, 5171k), warm light (3751lx, 4231k), and weak light (5lx, 5617k), and each lighting condition contains 50, 46, 45, 47, and 40 images, respectively. Besides, each captured image contains a face and a color checker.



(a) Acquisition Equipment



(b) Acquisition Environment

Figure 2. The lighting condition in a dark environment

We selected the D65 color checker as the target lighting environment for facial color correction. Considering the shooting angle in different lighting conditions, the color card data taken each time is slightly different. To avoid enlarging the difference in color correction results due to the difference in standard color card values, the average value of collected color card values in the D65 lighting condition is calculated as reference data.

## 2.2 Color Patch Selection

For the facial color correction task, the whole color patches are selected from color reference as training samples, which increases the complexity of the correction models (1). In addition, some colors of color patches in the ColorChecker (Figure 3) are unlikely to appear in the facial images. Inspired by the color patch selection method of Ref (16), we conducted group experiments on the selection of color patches in Xrite ColorChecker as shown below:

- (1) Group A: 8 basic skin-related color patches (numbered from 1 to 8): six different levels of grayscale patches and two skin color patches.
- (2) Group B: 12 face-related color patches (numbered from 1 to 12): six different levels of grayscale patches, 4 red-related color patches, and two skin color patches.
- (3) Group C: 24 standard color patches.



Figure 3. 24 Xrite ColorChecker (Numbered color patches)

### 2.3 Brightness Adjustment

The captured images in different lighting conditions will display different color information on the monitor. Therefore, we apply a white-balance-based algorithm (17) to detect and adjust the overall brightness of the captured images. The improved algorithm adjusts the overall color of white objects in the image according to their color drift. Since it is difficult to have pure white objects in facial images, the algorithm uses the color value of the white patch (No.6 patch in Figure 2) on the target color checker as the reference white pixels, instead of using the highest luminance value as the gain of the standard white.

In RGB color space, the original RGB value  $(R_{org}, G_{org}, B_{org})$  of the input image is obtained. Next, we transfer RGB color space to YCrCb color space and calculate the RGB average value  $(R_w, G_w, B_w)$  and the YCrCb average value  $(Y_w, Cr_w, Cb_w)$  of the No. 6 color patch in the target color space.

The first scale factor  $(R_{scale}, G_{scale}, B_{scale})$  using reference white pixels is done by the following formulas:

$$\begin{cases} R_{\text{scale}} = Y_w/R_w \\ G_{\text{scale}} = Y_w/G_w \\ B_{\text{scale}} = Y_w/B_w \end{cases} \quad (1)$$

Then the second scale factor  $(R_{\text{GWA}}, G_{\text{GWA}}, B_{\text{GWA}})$  is

$$\begin{cases} R_{\text{GWA}} = \text{Gray}/R_{\text{aver}} \\ G_{\text{GWA}} = \text{Gray}/G_{\text{aver}} \\ B_{\text{GWA}} = \text{Gray}/B_{\text{aver}} \end{cases} \quad (2)$$

where  $\text{Gray} = (R_{\text{aver}} + G_{\text{aver}} + B_{\text{aver}})/3$ , and  $(R_{\text{aver}}, G_{\text{aver}}, B_{\text{aver}})$  is the average value of  $(R_{\text{org}}, G_{\text{org}}, B_{\text{org}})$ .

## 2.4 Correction Regression Algorithm

The color card images captured under different illuminate environments are taken as data samples to obtain the target color values and the distorted color values of each color patch respectively. Then the color correction model learns the mapping relationship between the distorted color images and the target color images.

Suppose the target color checker has  $N$  color patches, and these three-channel values of the  $i$ -th color patch are  $R_{0i}$ ,  $G_{0i}$ , and  $B_{0i}$ . The three-channel values of the  $i$ -th color patch in the distorted color checker are  $R_i$ ,  $G_i$ , and  $B_i$  ( $i=1,2,3,\dots,N$ ). It can be represented as the following formula:

$$\begin{cases} R_{0i} = a_{11}s_{1i} + a_{12}s_{2i} + \dots + a_{1i}s_{mi} \\ G_{0i} = a_{21}s_{1i} + a_{22}s_{2i} + \dots + a_{2i}s_{mi} \\ B_{0i} = a_{31}s_{1i} + a_{32}s_{2i} + \dots + a_{3i}s_{mi} \end{cases} \quad (i=1,2,\dots,N) \quad (3)$$

where  $s_{mi}$  ( $m=1,\dots,M$ ) is formed by polynomials of  $R_i$ ,  $G_i$ , and  $B_i$  and the polynomial terms  $M = [1, R, G, B, R \times G, R \times B, G \times B, R^2, G^2, B^2]$ .

The above system of equations is transferred to the matrix as follows:

$$X = A^T * S \quad (4)$$

where  $X$  is the RGB matrix of a target color card with dimension  $3 \times I$ ,  $A$  is the conversion coefficient matrix to be solved with dimension  $I \times 3$ , and  $S$  is the polynomial regression matrix with dimension  $M \times I$ , which are respectively expressed as follows.

$$X = \begin{bmatrix} R_{01} & R_{02} & \dots & R_{0i} \\ G_{01} & G_{02} & \dots & G_{0i} \\ B_{01} & B_{02} & \dots & B_{0i} \end{bmatrix}, \quad A = \begin{bmatrix} a_{11} & a_{12} & a_{13} \\ a_{21} & a_{22} & a_{23} \\ \vdots & \vdots & \vdots \\ a_{s1} & a_{s2} & a_{s3} \end{bmatrix}, \quad S = \begin{bmatrix} s_{11} & s_{12} & s_{13} & \dots & s_{1I} \\ s_{21} & s_{22} & s_{23} & \dots & s_{2I} \\ s_{31} & s_{32} & s_{33} & \dots & s_{3I} \\ \vdots & \vdots & \vdots & \ddots & \vdots \\ s_{M1} & s_{M2} & s_{M3} & \dots & s_{MI} \end{bmatrix}$$

The regression matrix A obtained by the ridge regression algorithm is:

$$A = (MM^T + \lambda I)^{-1} MX^T \quad (5)$$

where I is the identity matrix and  $\lambda$  is the adjustment parameter to control the increased penalty force for  $\|\beta\|^2$ . The increase of  $\lambda$  can reduce the variance of model error and make the generalization ability stronger, but it will increase the deviation of model error. In the experiment, the generalized cross-validation method is employed in the ridge trace map to obtain the adjustment parameter  $\lambda$  that minimizes the mean square deviation of the model.

Finally, after obtaining coefficient regression matrix A, the corrected facial image is calculated by the following formulas:

$$X_{output} = A^T * S_{input} \quad (6)$$

where  $S_{input}$  is the cropped image after 68-point face detection, and  $X_{output}$  is the RGB matrix of the corrected facial image. Finally, the corrected facial image can be calculated to realize the color correction.

### 3 Experimental Results and Analysis

#### 3.1 Experimental Settings

Because of physiology, psychology, and chromaticity, the evaluation of color correction algorithms is very challenging (18). The evaluation methods have two components: objective evaluation and subjective evaluation.

The objective evaluation is to calculate the chromatic distance of color patches. If the chromatic distance  $\Delta E \geq 3.0$  in the CIE LAB color space, two different visual perceptions will appear. The value is calculated by the following formulas:

$$\Delta E_k = \sqrt{(L_{ki} - L_{kj})^2 + (a_{ki} - a_{kj})^2 + (b_{ki} - b_{kj})^2} \quad (7)$$

where k represents the k-th color patch, i is the distorted color checkers captured in different lighting conditions, and j denotes the standard color checker in the D65 lighting condition.  $\Delta E_{mean}$  is the average color difference of the color patches.

The subjective evaluation is that a certain number of observers visually evaluate the color quality of the corrected images. From the subjective perception of Figure 4, the color checkers under different lighting conditions have different visual perceptions and the color checkers after correction are similar to that under the D65 illumination condition.




Figure 4. The captured color checkers (upper row) and their corrected results of our method (bottom row)

## 3.2 RESULTS AND DISCUSSION

### 3.2.1 Results of Baseline Method Selection

To obtain the evaluation results for the objective analysis, we selected color patches ( $10 \times 10$  pixels) from each color checker to calculate the chromatic distance. We compared POLY-(1), SVR-(11), and CNN-based(18) color correction methods with 24 color patches. We used SVR (11) and Gaussian kernels to build the mapping relationship, in which there are only three parameters needed to be optimized, i.e., penalty factor C, the gamma of Gaussian kernel g, and epsilon of lost function p. Eventually, C is set to 20, g is set to 0.03 and p is set to 0.

As can be seen from Figure 5, the  $\Delta E_{\text{mean}}$  of above color correction algorithms based on 24 color patches under five illumination conditions were significantly reduced, and even the  $\Delta E_{\text{mean}}$  of CNN algorithm was below 2. However, the visual comparisons of CNN (see Figure 5-6) showed significant distortion, demonstrating that CNN was not suitable for small datasets. Moreover, no significant difference was noted between  $\Delta E_{\text{mean}}$  of uncorrected images and that of SVR. Moreover, the correction results of POLY were better than those of SVR. Therefore, we applied the polynomial regression-based algorithm for subsequent experiments on color patch selection.




Figure 5.  $\Delta E_{\text{mean}}$  of 24 color patches under five different lighting conditions. From left to right: D50, D65, D65warm, warm and weak. Black lines depict significant difference between  $\Delta E_{\text{mean}}$  of three methods and the uncorrected images




Figure 6. Visual comparisons of three methods for facial color correction. From top to bottom: uncorrected images, and corrected images by POLY, SVR, and CNN, respectively.

### 3.2.2 Results of Color Patches Selection

According to Section 2.1, we compared the polynomial-based algorithm on the above three groups (Group A, Group B, and Group C). As shown in Figure 7, the values of group A are higher than the color difference before correction. On the contrary, the average difference between groups B and C was similar, and both were lower than the color difference before correction. Figure 8 shows examples of facial color correction for visual comparison. The figure shows that the difference between the corrected face images of groups B and C is not significant.

It can be seen that it is not essential to select all color patches (Group C) as training samples to correct facial images. In addition, irrelevant color patches will affect the results of the objective analysis. Therefore, it is demonstrated that the 12 selected color patches are sufficient for this facial color correction task.




Figure 7.  $\Delta E_{\text{mean}}$  of three groups under five different lighting conditions. From top to bottom: Groups A, B, and C. respectively. From left to right: D50, D65, D65warm, warm and weak. Different color Discrete points present the chromaticity distances of each color patch in different groups.




Figure 8 Sample results of facial color correction on Groups A, B, and C. From top to bottom: Groups A, B, and C. respectively. From left to right: D50, D65, D65warm, warm and weak.

### 3.2.3 Results of Color Correction Algorithm Selection

We compared the proposed method with the polynomial-based correction method (3) (19). The results of the objective evaluation are shown in Figure 9. In the three methods, our proposed algorithm significantly outperforms other methods in the objective evaluation. Especially, the  $\Delta E_{\text{mean}}$  of three methods in the weak illuminate environment have a significant difference. Figure 8f showed that the  $\Delta E_{\text{mean}}$  of our method under D50 and weak did not have significance difference with that of D65. Moreover, the  $\Delta E$  of our method in five different lighting conditions are lower than 3.0 (Figure 9f), which shows that our method is suitable for facial image color correction task.




Figure 9.  $\Delta E_{\text{mean}}$  of 12 color patches of three methods. (a)-(e)  $\Delta E_{\text{mean}}$  of three methods under five different lighting conditions. (f)  $\Delta E_{\text{mean}}$  of our method under five different lighting conditions.

Meanwhile, Figure 10 shows the qualitative performance of different color correction algorithms in five different lighting conditions. From this subjective vision. From the subjective perception, before correction under warm and D65warm lighting conditions, the color tone of the facial image is yellowish. After correction, the color tone becomes ruddy, but the degree of ruddy by different algorithms varies. It is obvious that the facial image after correction by other algorithms has the most reddish-dark tone under the weak lighting condition. The performance of our method is more consistent with the observed colors by human beings. All these results demonstrate the

effectiveness of our proposed color correction algorithm.




Figure 10. Sample results of different color correction algorithms. From top to bottom: uncorrected images, corrected images by Ref (3), corrected images by Ref (19), and corrected images by our method. From left to right: D50, D65, D65Warm, Warm, and Weak lighting conditions, respectively.

### 3.2.3 Ablation Studies with Brightness Adjustment

To further evaluate the effectiveness of our proposed method, we conducted ablation studies with brightness adjustment. The results were shown in Figures 11 and 12. In addition, we applied brightness adjustment as preprocessing in Ref (3) and Ref (19). From Figure 11, we find that using only brightness adjustment did not satisfy the needs of color correction tasks. When we respectively appended the brightness adjustment operation to Ref (3) and Ref (19), the performance was slightly improved. Moreover, the objective analysis shows that the improvement is especially noticeable in the weak lighting condition.




Figure 11.  $\Delta E_{\text{mean}}$  of seven methods under five different lighting conditions.

Meanwhile, we can see that the facial images captured in the different lighting conditions have different highlight areas because these images suffered from the dynamic lighting variety. For example, in the fifth column of images in Figure 12, the forehead area in the weak lighting condition obviously has a highlight area. When we employed the brightness adjustment as the first step, it produced a noticeable change in this area. Though the  $\Delta E_{\text{mean}}$  of Ref (19) + brightness adjustment was reduced, the color tone of these facial images is still yellowish. Besides, the color tone of the facial image of Ref (19) + brightness adjustment is more reddish, especially the facial image captured in D50 lighting condition. From the visual comparison, the corrected facial images of our method are better than the performance of the other methods.




Figure 12. Sample results of different color correction algorithms.

### 3.3 Discussion

In this section, we discuss the performance of the proposed color correction method compared to other methods in different lighting conditions. To obtain satisfactory perceptual adaptation, we proposed a two-stage color correction algorithm for facial images.

With the development of deep learning, many scholars have used convolutional neural networks to obtain satisfactory performances, but this network requires a large number of images, which is an expensive and time-consuming process. For example, in

---

Figures 4 and 5, CNN can achieve good results of the chromatic distances, but it cannot obtain a satisfactory subjective appearance. Therefore, we need to analyze both objective evaluation and subjective evaluation for the facial image color correction task. Additionally, SVR is a regression algorithm for small sample datasets, but it can be seen from Figure 5 that the POLY is more suitable for facial image color correction.

Previous methods for color correction of the facial image have focused on single-stage algorithms. However, these methods ignore adjusting brightness under different lighting conditions. We utilized the adaptive white balance algorithm as the first step to adjust the brightness and then applied the ridge regression algorithm to optimize the color correction model (see Figure S1). Due to the features of facial shape, the facial image contains forehead highlight areas and even different highlight areas under different lighting conditions.

Meanwhile, as can be seen from Figure 8f, the chromaticity distances after correction in different illumination environments did not achieve consistency with those in the D65 environment. This difference may be caused by the illumination and color temperature of the environment.

Besides, we selected 12 face-related color patches to optimize the color correction model. As shown in Figure 6, the experimental results demonstrated that the  $\Delta E_{\text{mean}}$  of Groups B and C performed very close. Therefore, 12 facial-related color patches are sufficient for the facial image color correction task, which is beneficial for reducing computational complexity and time.

Even though the proposed network has achieved good results under different lighting conditions, it still has some limitations: (1) The dataset only contains Asian and the selected color patches mainly incline to Asian skin color. (2) Our model is designed for five different lighting conditions. In recent years, the medical facial image captured in an open environment has become increasingly desirable for various medical diagnosis tasks.

In our future work, we aim to design a different race facial image color correction in an open environment. The dataset should contain more different kinds of skin colors. We hope that more teams will explore this research direction, which is of great importance for clinical diagnosis.

#### **4 Conclusion**

This study proposed a two-stage color correction algorithm for the facial image. The experimental results show that 12 color patches for color correction of face distortion images are sufficient for this task. It is also further demonstrated that the regression model achieves smaller chromaticity differences in different lighting conditions. Moreover, we conducted the ablation experiments for the white balance algorithm, and the experimental results show that the color difference value error can be better reduced. However, some limitations must be noted. The proposed method must still be improved considering the open environment for the facial color correction task. Accordingly, we need to design a facial color correction in an open environment in the future.

---

## CONSENT

All authors declare that written informed consent was obtained from the patient for publication of this paper and accompanying impaperages. A copy of the written consent is available for review by the Editorial office/Chief Editor/Editorial Board members of this journal.

## ETHICAL APPROVAL

All authors hereby declare that all experiments have been examined and approved by the appropriate ethics committee and have therefore been performed in accordance with the ethical standards laid down in the 1964 Declaration of Helsinki.

## REFERENCE

1. Niu JL, Zhao CB, Li GZ. A comprehensive study on color correction for medical facial images. *Int J Mach Learn Cybern* [Internet]. 2019;10(5):935–47. Available from: <http://dx.doi.org/10.1007/s13042-017-0773-6>
2. Lu Y, Li X, Gong Z, Zhuo L, Zhang H. TDCCN: A two-phase deep color correction network for Traditional Chinese Medicine tongue images. *Appl Sci*. 2020;10(5):1–21.
3. Finlayson GD, Darrodi MM, Mackiewicz M. The Alternating Least Squares Technique for Nonuniform Intensity Color Correction. *Color Res Appl*. 2015;40(3):232–42.
4. Finlayson GD, Mackiewicz M, Hurlbert A. Color Correction Using Root-Polynomial Regression. *IEEE Trans Image Process*. 2015;24(5):1460–70.
5. Zhang B, Wang X, Karray F, Yang Z, Zhang D. Computerized facial diagnosis using both color and texture features. *Inf Sci (Ny)* [Internet]. 2013;221:49–59. Available from: <http://dx.doi.org/10.1016/j.ins.2012.09.011>
6. Zhuo L, Zhang P, Qu P, Peng Y, Zhang J, Li X. A K-PLSR-based color correction method for TCM tongue images under different illumination conditions. *Neurocomputing*. 2016;174(8):815–21.
7. Yang B, Chou HY, Yang TH. Color reproduction method by support vector regression for color computer vision. *Optik (Stuttg)* [Internet]. 2013;124(22):5649–56. Available from: <http://dx.doi.org/10.1016/j.ijleo.2013.04.036>
8. Ding X, Jiang Y, Qin X, Chen Y, Zhang W, Qi L. Reading face, reading health: Exploring face reading technologies for everyday health. In: *Conference on Human Factors in Computing Systems*. 2019. p. 1–13.
9. Zhuo L, Zhang J, Dong P, Zhao Y, Peng B. An SA-GA-BP neural network-based color correction algorithm for TCM tongue images. *Neurocomputing* [Internet]. 2014;134:111–6. Available from: <http://dx.doi.org/10.1016/j.neucom.2012.12.080>
10. Zhang J, Yang Y, Zhang J. A MEC-BP-Adaboost neural network-based color correction algorithm for color image acquisition equipments. *Optik (Stuttg)* [Internet]. 2016;127(2):776–80. Available from: <http://dx.doi.org/10.1016/j.ijleo.2015.10.120>
11. Lixia M. Color Correction of Face Image and its Application in

Hepatopathy Diagnosis. Harbin Institute of Technology; 2009.

12. Akazawa T, Kinoshita Y, Shiota S, Kiya H. Three-color balancing for color constancy correction. *J Imaging*. 2021;7(10).
13. Yan X, Wang G, Jiang G, Wang Y, Mi Z, Fu X. A natural-based fusion strategy for underwater image enhancement. *Multimed Tools Appl*. 2022;81(21):30051–68.
14. Huang H, Liao N, Zhao C, Fan Q. White Balance Conversion Method of Different Camera Based on Triangle Affine Transform. *Lect Notes Electr Eng*. 2022;896 LNEE:1–7.
15. Viola P, Jones M. Rapid object detection using a boosted cascade of simple features. In: *Proceedings of the 2001 IEEE Computer Society Conference on Computer Vision and Pattern Recognition(CVPR 2001)*. 2001. p. 511–8.
16. Zhang D, Zhang H, Zhang B. *Tongue Image Analysis*. Singapore:Springer; 2017.
17. Lam EY. Combining gray world and retinex theory for automatic white balance in digital photography. In: *Proceedings of the Ninth International Symposium on Consumer Electronics, 2005 (ISCE 2005)*. 2005. p. 134–9.
18. Lu Y, Li X, Zhuo L, Zhang J, Zhang H. Dccn: A deep-color correction network for traditional Chinese medicine tongue images. *2018 IEEE Int Conf Multimed Expo Work ICMEW 2018*. 2018;1–6.
19. Su Q, Cheng H, Sun W, Zhang F. A Novel Correction Algorithm Based on Polynomial and TPS Models. In: *2011 International Conference of Information Technology, Computer Engineering and Management Sciences*. 2011. p. 52–5.

## APPENDIX




Figure S1. Sample results of different color correction algorithms. From top to bottom: uncorrected images, corrected images by Ref (3), corrected images by Ref (19), and corrected images by our method. From left to right: (a)D50, (b)D65, (c)D65warm, (d)warm, and (e)weak lighting conditions, respectively.