

Original Research Article

Factors affecting validity of PfHRP2-based rapid diagnostic test kit in the diagnosis of malaria in Afikpo, Ebonyi State: a pilot study.

ABSTRACT

Aim: To assess the factors affecting validity of histidine rich protein II (HRP2)-based rapid diagnostic test (RDT) kit for diagnosis of malaria in Afikpo North Local Government Area, Ebonyi State.

Study design: Quantitative method using both structured questionnaires and serological tests were employed to collect data on factors affecting validity of HRP2-based RDT kits results for malaria diagnosis.

Place and duration of study sample: Four different health facilities including Akanu Ibiam Federal Polytechnic Medical Centre, Unwana, between November 2022 and March 2023.

Methodology: We interviewed eleven (11) health workers (HWs) (5 males and six females) on factors affecting validity of RDT kits. We also included 50 patients (19 males and 31 females; aged 1-40 years) with symptoms of malaria to validate some of the factors mentioned by the health workers. Replicates of CareStart[®]™ malaria RDT kits were used to test some of the factors on blood samples collected from the patients. Data collected were analyzed using Chi-square test.

Result: Majority of the HWs use microscopy (54.5%) while 45.5% use RDT for diagnosis of malaria. Majority of the HWs claimed that quantity of buffer reagent (50%) and wait time (14.3%) affects performance of RDT kits. However, there was no statistically significant difference ($p > 0.05$) in the prevalence of malaria (36%) among the different buffer quantities and wait times tested. Hence, the tested buffer reagent volumes and wait times has no effect on accuracy of rapid diagnostic test. It is important to validate these findings in future studies with larger sample size.

Key words: malaria, rapid diagnostic test kit, buffer reagent, wait time.

INTRODUCTION

Nigeria accounts for the highest proportion (27%) of global malaria cases [1]. Pregnant women and children under five years old are most affected. Most malaria-related deaths happen within few days from the time an individual start showing symptoms, and when early diagnosis and prompt treatment with an effective antimalarial drug is delayed. Hence, prompt and accurate laboratory diagnosis is important for both rapid and effective management and surveillance of malaria [2]. Diagnosis of malaria using light microscopy and polymerase chain reaction (PCR) methods are highly recommended due to their high sensitivity and specificity, but specific challenges have limited their use in malaria endemic countries like Nigeria. Fortunately, the introduction of rapid diagnostic test (RDT) device had made these challenges surmountable [3-6].

Malaria rapid diagnostic test (mRDT) kit is a device used to confirm malaria infection in patients presenting symptoms. It detects specific *Plasmodium* antigen in small quantity of fresh blood

using lateral flow immunochromatography [7]. The types of mRDT in use can detect one or combination of the following antigens: (1) Histidine-rich protein 2 (HRP2)- a water soluble protein -produced by trophozoites and gametocytes of *P. falciparum* only; (2) *Plasmodium* lactate dehydrogenase (pLDH) enzyme produced by all *Plasmodium* species infective to human; and (3) *Plasmodium* aldolase enzyme produced by all species of human *Plasmodium* parasites [8].

Although, mRDT is recommended for malaria diagnosis in resource limited areas that are endemic for malaria, studies had reported varying levels of its usage and adherence to test results among health workers (HWs) [9]. Report had shown that RDT was not used by health workers in Enugu state because of the pervasive notion that RDT results were inaccurate [10]. A previous study had also shown that 37.6% of health workers in Ebonyi State had poor perception of malaria RDT kits [9]. Low uptake and confidence in rapid diagnostic test kits may result to inappropriate prescription of antimalarial drugs, economic waste and delayed detection of the primary cause of malaria-like symptoms.

Hence, this pilot study is aimed at assessing factors affecting the validity of RDT kit results in the diagnosis of malaria in the study area.

MATERIAL AND METHODS

Study Area

The study was conducted in Afikpo under Afikpo North Local Government Area (L.G.A.) of Ebonyi State, Nigeria. Ebonyi state has the highest prevalence of malaria in southeast Nigeria [11] and malaria is the most prevalent medical condition treated in healthcare facilities across the state [12, 9].

Afikpo is located between latitudes 5° 4' N and 6° 3' N and longitudes 7° 5' E and 7° 55'E. There are predominantly two main seasons: rainy season between April – October and dry season between November – March. The annual rainfall is about 160mm – 220mm with maximum precipitation occurring between July and September. Atmospheric temperature is between 23.4°C and 29.9°C and relative humidity is between 60 – 80%. Afikpo is bounded in the North by Ohazara L.G.A., Ebonyi State, in the East by Afikpo South L.G.A., in the South by Cross River State, and in the West by Abia State. The population of people in Afikpo is about 672,000 persons [13-15].

Administration Of Questionnaires

Well-structured questionnaires were administered to medical laboratory scientists working in hospitals and primary health centres within three communities (Afikpo, Enohia, and Unwana) within Afikpo North L.G.A. The questionnaires contained questions for collecting data on sex, age, level of education, marital status, knowledge on causes of malaria, mode of malaria transmission, diagnosis of malaria, and perception on factors affecting validity of RDT kits.

Sampling of malaria symptomatic patients

Blood samples were collected from 50 malaria patients at the polytechnic medical centre. The blood samples were those for malaria by the medical lab scientist at the polytechnic medical centre.

Procedure for malaria rapid diagnostic test

The expiry date of the packet was checked, the gloves were worn. The packet was opened. Three cassettes were placed on a clean flat surface and labelled clearly with the patient's ID number. The loop was used to put the drop of blood into the square hole marked "S" and the loop was discarded. Two drops of buffer were put into the round hole marked "A", for low buffer volume 1 drop of buffer was put, for high buffer volume 4 drops of buffer were put. After the buffer was added, I waited for 20 minutes to read test results then results were recorded.

How the RDT result was interpreted

- i. One red line in window "C" and one red line in window "T" means that the patient was positive for malaria.
- ii. One red line in window "C" and no line in window "T" means that the patient was negative for malaria.
- iii. No line in window "C" means the test was invalid.
- iv. A line in window "T" and no line in window "C" also means the test was damaged. Results were invalid. Another test was conducted if result was invalid.

Testing the effect of varying buffer volume and wait time on accuracy of malaria RDT tests.

Carestart™ RDT was used for this study.

3 replicates of RDT were used to conduct test on each blood sample as shown below

1 RDT for low buffer volume with 1 drop

1 RDT for high buffer volume with 4 drops

1 RDT for recommended buffer volume (2 drops) by manufacturer (control).

Result of the test was recorded and compared to microscopy test result.

Data analysis

Data collected was entered in SPSS software version 16.0 for analysis. Chi-square test and Fisher's test were used to determine the effect of buffer volume on accuracy of malaria test results. The result was presented using percentages in standard tables.

RESULTS AND DISCUSSION

Demographic characteristics of health workers (HWs) interviewed

Data in Table 1 showed that 45.5% of the MLSs interviewed were male while 54.5% were female. Majority of the medical laboratory scientist are within the age bracket of 31-40 years (72.7%). They all have tertiary education (100%). Majority of them (63.3%) have 6-10 years work experience and worked in the hospital (45.5%).

Demographic characteristics of malaria symptomatic patients.

A higher proportion (62%) of patients diagnosed for malaria were female while 38% were male. In terms of their age, majority of the patients (64%) are within the age bracket of 21-30 years (Table 2).

Usage and factors affecting validity of RDT according to the health workers (HWs).

Result in Table 3 shows that 54.5% of MLS use microscopy while 45.5% use RDT to diagnose malaria in their laboratories. With respect to whether RDT is more accurate than microscopy, majority (81.8%) responded "no" while 18.2% said "yes". With respect to the factors that may affect RDT results, 54.5% believed that buffer quantity can affect RDT result while others mentioned wait time (9.1%), wrong placement of buffer and blood (9.1%), climatic conditions (7.1%), and expired RDT kit (7.1%). However, 18.2% have no idea of any factor that can affect RDT result.

Table 1: Demographic characteristics of health workers

(HWs) interviewed

	n(%)
Sex	
Male	5 (45.5)
Female	6 (54.5)
Age (years)	
20-30	1 (9.1)
31-40	8 (72.7)
41-50	1 (9.1)
>50	1 (9.1)

Level of education

Secondary	0 (0)
Tertiary	11 (100)

Years of work experience

1-5	1 (9.1)
6-10	7 (63.6)
11-15	1 (9.1)
16-20	0 (0)
>20	2 (18.2)

Type of healthcare

Primary Health Centre	4 (36.4)
Medical Centre	1 (9.1)
Hospital	5 (45.5)
Private Medical Lab.	1 (9.1)
Total	11

Table 2: Demographic characteristics of patients diagnosed for malaria

	n(%)
Sex	
Male	19 (38)
Female	31 (62)
Total	50 (100)
Age (years)	

0-10	5 (10)
11-20	8 (16)
21-30	32 (64)
31-40	5 (10)
Total	50 (100)

Table 3: Health workers' usage and factors affecting validity of RDT kits.

	n(%)
How MLS diagnose malaria in their lab.	
Use RDT	5 (45.5)
Use microscopy	6 (54.5)
Total	11
Is RDT more accurate than microscopy?	
Yes	2 (18.2)
No	9 (81.8)
Total	11
Perceived factors that affect RDT result	
Wait time	2 (14.3)

Buffer quantity	7 (50.0)
Wrong placement of buffer and blood	1 (7.1)
Climatic conditions	1(7.1)
Expired RDT	1(7.1)
No idea	2 (14.3)
Total	14

Effect of buffer quantity and wait time on RDT result

There was no significant difference ($p=1$) in the number of samples that tested positive and negative within the various test groups. Out of the 50 samples analyzed, 36.0% were positive to *Plasmodium falciparum* while 64.0% were negative. With respect to quantities of buffer volumes tested, the prevalence of malaria (36%) was the same at low, normal, and high buffer volumes respectively. With respect to the various wait times tested, the prevalence of malaria (36%) was also the same at low, normal, and high wait times respectively.

Table 4: Effect of buffer quantity and wait time on validity of RDT kits used for diagnosis of malaria in symptomatic patients.

Test groups	RDT result n(%)		Total	Fisher's test
	Positive	Negative		
Quantity of buffer				
Low	18 (36.0)	32 (64.0)	50 (100)	Z=0

Normal (recommended)	18 (36.0)	32 (64.0)	50 (100)	P=1
High	18 (36.0)	32 (64.0)	50 (100)	
Wait time				
Low	18 (36.0)	32 (64.0)	50 (100)	
Normal (recommended)	18 (36.0)	32 (64.0)	50 (100)	
High	18 (36.0)	32 (64.0)	50 (100)	

Discussion

The accuracy and validity of malaria rapid diagnostic test (RDT) kits can be influenced by various factors, and temperature, humidity and RDT preparation are among such factors [16]. The findings from this study had shown that majority of the health workers had a wrong perception of factors that can affect the validity of RDT kits in the diagnosis of malaria. There is no valid report that reagent (buffer) volume and wait time affected the performance of RDT kits. Going a step further in this study, we reported that the various volumes of buffer solution and wait time assessed did not affect the validity of RDT kit performance.

It is interesting to note that the use of inappropriate buffer reagents can affect the performance of RDT kits. Buffer reagents play a crucial role in ensuring the proper functioning of RDTs by facilitating the reaction between the sample and the test components. In the case of malaria RDTs, the buffer solution is designed to optimize the conditions for the interaction between the target antigens in the patient's blood and the antibodies or other detection molecules on the test strip. Substitution of buffer reagents can affect the performance of RDT kits in several ways including its sensitivity and specificity, reaction rate on the test strip, interference with test components, and storage stability [17, 18]. However, in the present study, substitution of buffer reagent was not investigated.

Hence, the use of buffer reagents recommended by the RDT kit manufacturer is encouraged to ensure optimal performance. The present does not in any way encourage RDT kit users to deviate from manufacturers' specific instructions on the preparation and use of the buffer solution. It is recommended that users of RDT kits refer to the specific product documentation and guidelines provided by the RDT kit manufacturer for accurate information on buffer reagents and test procedures.

Conclusions

The various buffer reagent volumes and wait times tested in the diagnosis of malaria using RDT device, did not affect the validity of the RDT kits.

Consent

All authors declare that oral consent was obtained from the patients that participated in the study before sample collection.

Ethical Approval: the study proposal was reviewed by the Biology research unit of Science Laboratory Technology department, A.I.F.P.U., Ebonyi State. Approval to conduct the study was also obtained at the Medical Centre of the host institution. Oral consent was sought from study participants before blood sample collection.

References

1. WHO. World malaria report 2021. Geneva, World Health Organization. 2021. <https://www.who.int/teams/global-malaria-programme/reports/world-malaria-report-2021>. Accessed 07 Dec 2021
2. WHO. Malaria: diagnostic testing; microscopy. Geneva, World Health Organization. 2018. <https://www.who.int/malaria/areas/treatment/overview/en/>. Accessed 18 July 2019
3. World Health Organization (2012). Foundation for Innovative New Diagnostics, TDR, Roll Back Malaria. Malaria Rapid Diagnostic Tests. Available from: <http://www2.wpro.who.int/sites/rdt/home.htm>
4. Lal S, Ndyomugenyi R, Paintain L, Alexander ND, Hansen KS, Magnussen P, et al. (2016). Community health workers adherence to referral guidelines: evidence from studies introducing RDTs in two malaria transmission settings in Uganda. *Malar J. BioMed Central*; 2016; 1–13. <https://doi.org/10.1186/s12936-015-1044-1>
5. Kyabayinze DJ, Asiimwe C, Nakanjako D, Nabakooza J, Counihan H, Tibenderana JK. Use of RDTs to improve malaria diagnosis and fever case management at primary health care facilities in Uganda. *Malar J* [Internet]. 2010; 9:200. Available from: <http://malariajournal.biomedcentral.com/articles/10.1186/1475-2875-9-200> PMID: 20624312
6. Chandler CIR, Whitty CJM, Ansa EK. How can malaria rapid diagnostic tests achieve their potential? A qualitative study of a trial at health facilities in Ghana. *Malar. J.*; 2010; 9:95. Available from: <https://doi.org/10.1186/1475-2875-9-95> PMID: 20398262
7. Odaga J, Sinclair D, Lokong JA, Donegan S, Hopkins H, Garner P. Rapid diagnostic tests versus clinical diagnosis for managing people with fever in malaria endemic settings. *Cochrane Database Syst Rev.*; 2014; <https://doi.org/10.1002/14651858.CD008998.pub2>.
8. Ghana Health Service (2019). Rapid Diagnostic Test (RDT) Kit. www.ghanahealthservice.org/malaria/item.php?nmcpiid=72&nmcpscicid=127&nmcpcid=90 (Retrieved 31st May, 2019).
9. Obi IF, Sabitu K, Olorukooba A, Adebowale AS, Usman R, Nwokoro U, et al. (2019) Health workers' perception of malaria rapid diagnostic test and factors influencing compliance with test results in Ebonyi state, Nigeria. *PLoS ONE* 14(10): e0223869. <https://doi.org/10.1371/journal.pone.0223869>
10. Mokuolu OA, Ajumobi OO, Ntadom GN, Adedoyin OT, Roberts AA, Agomo CO, Edozieh KU, Okafor HU, Wammanda RD, Odey FA, Maikore IK, Abikoye OO, Alabi AD, Amajoh C, and Audu BM. Provider and patient perception of rapid diagnostic test use in Nigeria: a cross-

sectional evaluation. *Malaria Journal*, 2018; 17:200 <https://doi.org/10.1186/s12936-018-2346-x>

11. National Malaria Elimination Programme (NMEP), National Population Commission (NPopC), National Bureau of Statistics (NBS) and II. Nigeria Malaria Indicator Survey 2015. Abuja, Nigeria and Rockville, Maryland, USA; 2016.
12. Odoh B. (2015). Ebonyi State Factsheet. Abakaliki.
13. Iloeje PU. Regional Geography for West Africa. 2nd Ed. Macmillan Publishers, Nigeria; 1981.
14. National Population Commission (NPC) (2006). Census and demographic Report of Nigeria. Published by the Nigerian Minting and Printing Press Abuja Nigeria.18-45.
15. Ngele KK, Alaribe AAA, Shoniron OT. Using clinical signs to determine the endemicity level of lymphatic filariasis in Afikpo North L.G.A. Ebonyi State, Nigeria. *Global Journal of Bio-Science and Biotechnology*, 2013; vol. 2(4):305-310.
16. Bell D, Peeling R. Evaluation of rapid diagnostic tests: malaria. *Nat Rev Microbiol*; 2006; 4 (Suppl 9), S34–S38. <https://doi.org/10.1038/nrmicro1524>
17. World Health Organization (WHO). (2019). Malaria rapid diagnostic test performance: Results of WHO product testing of malaria RDTs: Round 8 (2016-2018). Geneva: World Health Organization.
18. Gillet P, Mori M, Van Den Ende, J. Buffer substitution in malaria rapid diagnostic tests causes false-positive results. *Malaria Journal*, 2010; 9(1), 215.