

Original Research Article

Correlation of Volume Computed Tomography Dose Index and Dose Length Product with Acquisition Parameters in Abdomino-Pelvic and Cranial Computed Tomography Imaging

ABSTRACT

Introduction: The use of Computed Tomography (CT) in medical diagnosis showing more complex fractures and other pathological findings in 3D, delivers relatively higher radiation doses to patients when compared to conventional X-rays. In order to maintain good diagnosis while reducing patients' dose, it is necessary to assess the contributions of CT scan acquisition parameters (kV, mAs and scan length) to patient dose, and improve the selection of those parameters. This work aimed at investigating the relationship between the Volume CT Dose Index (CTDI_{vol}) and Dose Length Product (DLP) with scan acquisition parameters in order to establish the Correlation of CTDI_{vol} and DLP with scan acquisition parameters in Abdomino-Pelvic and Cranial CT.

Methodology: A total of 61 existing patients' data with scan acquisition parameters were collected: Cranial CT (39) and Abdomino-pelvic CT (22). Pearson correlation analysis was used to establish the relationship between scan acquisition parameters and Total CTDI_{vol} and Total DLP.

Results: The Total CTDI_{vol} ranged from 135.6 to 215.5 mGy for cranial CT and 26.1 to 124.4 mGy for Abdomino-pelvic, while DLP ranged from 2631.0 to 6542.8 mGy.cm for cranial and 1147.5 to 5441.8 mGy.cm for Abdomino-pelvic.

The CTDI_{vol} showed moderate positive correlation with kV ($r = 0.625$, P value < 0.01) for Cranial and strong positive correlation ($r = 0.821$, P value < 0.01) for Abdomino-pelvic CT, while DLP showed strong positive correlation ($r = 0.767$ and 0.839 , P value < 0.01) for both Cranial and Abdomino-pelvic. The CTDI_{vol} and DLP showed strong positive correlation with mAs ($r = 0.843$ and 0.839 , P value < 0.01) for both cranial and abdomino-pelvic CT. CTDI_{vol} and DLP showed a strong positive correlation with scan length ($r = 0.712$ and 0.881 , P value < 0.001) for both Cranial and Abdomino-pelvic CT. Increase in kV, mAs and scan length may cause significant increase in CTDI_{vol} and DLP in Cranial and Abdomino-pelvic CT.

Conclusion: The relationships between CTDI_{vol} and DLP with scan acquisition parameters in adult patients have been investigated for both Cranial CT and Abdomino-pelvic CT. It was observed that scan acquisition parameters (kV, mAs and Scan length) had linear relationship with CTDI_{vol} and DLP in both investigations.

Keywords: [Computed Tomography, Correlation, Volume Computed Tomography Dose Index (CTDI_{vol}), Dose Length Product (DLP), Abdomino-pelvic CT Scans.]

1. INTRODUCTION

Computed Tomography (CT) is an efficient diagnostic tool for the diagnosis and management of various clinical indications because of its ability to acquire high-quality 3-D diagnostic images. CT scans are particularly beneficial in the diagnosis of Cranial and Abdominal ailments because of its effectiveness in proper diagnosis. Therefore, it enables faster and accurate diagnosis without recourse surgical procedure. The amount of radiation dose associated with CT is 10 – 100 times greater than conventional X-ray which puts patients to higher risk of cancer induction (Smith-Bindman, 2008).

Computed Tomography Dose Index (CTDI) and Dose Length Product (DLP) are considered as dose descriptors in CT. They are displayed at the end of each scanned protocol on the summary page of the scan monitor. The $CTDI_{vol}$ is a standardized measure of the radiation output of a CT scanner although not a direct **measure** of patient dose. It is the amount of radiation delivered to one slice of the body part being examined over a period of time in the helical CT scanner. The DLP is the total radiation energy delivered to all the slices of the body part being scanned during a particular protocol (Huda, 2011).

The CTDI and DLP depend on the following acquisition parameters; tube loading (mAs), kilovoltage (kV), body size and scan length. The radiation exposures to patients undergoing CT examinations are determined by two factors: equipment-related factors, which have to do with the design of the scanner with respect to dose efficiency, and application-related factors, which have to do with the selection of protocols and parameters by the radiographer (Yu et al., 2009).

Studies emphasizing on CT parameters and radiation dose have been performed in the past. (Bjorkdahl and Nyma, 2010; Christner et al., 2012; Israel et al., 2010; Zarb and McEntee, 2010). However, a systematic analysis of the relationship between CT dose descriptors (CTDI and DLP) of Abdomino-pelvic and Cranial CT with varied parameters (kV, mAs, weight and scan length) has not been sufficiently investigated. In order to investigate how variations in acquisition parameters affects patients' dose, the relationship between the acquisition parameters and $CTDI_{vol}$ and DLP needs to be evaluated (Aweda and Arogundade, 2007). The rapid evolution of CT technology and its application in patients' diagnosis, requires detailed understanding and critical investigation.

This study focuses on how scan input parameters influence patient dose during CT investigations, using Cranial and Abdomino-pelvic CT procedures. The findings will be useful in moderating the acquisition parameters (kV, mAs, weight and scan length) in relation to the scanner output.

2. Theoretical Background

Two related measures of CT radiation dose are available on CT consoles: the Volume Computed Tomography Dose Index ($CTDI_{vol}$) and the Dose Length Product DLP. Both functions have reasonable proxies for absorbed dose but do not represent the actual patient dose.

Computed Tomography Dose Index (CTDI) represents the radiation dose of a single CT slice and is determined using acrylic phantoms. These acrylic phantoms are cylinders of a standard length and are generally in diameters of 16 cm and 32cm. The original incarnation of the CTDI was based on an axial CT scanner. Several variations of the CTDI have been defined. The $CTDI_{100}$ is the dose contribution from a 100 mm range, centered on the index slice. The weighted CTDI ($CTDI_w$) is the weighted sum of two-thirds peripheral dose and one third central dose in a 100 mm range in acrylic phantoms (Goldman, 2007, **Annam et al., 2019**).

The Volume Computed Tomography Dose Index ($CTDI_{vol}$) represents the scanner output. It is defined as $CTDI_w$ divided by the beam pitch factor (equation 1). It is the most commonly used index for modern Multiple Detector Computed Tomography (MDCT) equipment.

The $CTDI_{vol}$ can be calculated as:

$$CDDI_{vol} = \frac{NT}{I} (CTDI_w) \quad (1)$$

Where: $CTDI_w$ is the weighted or average CTDI given across the field of view, N is the number of simultaneous axial scans per x-ray source rotation, T is the thickness of one axial scan (mm) and I is the table increment per axial scan (mm).

In helical CT, the ratio of I to NT is the pitch. Therefore, in helical mode,

$$CTDI_{vol} = \frac{1}{Pitch} * CTDI_w \quad (2)$$

$CTDI_{vol}$ represents the dose for a specific scan protocol which takes into account gaps and overlaps between the radiation dose profile from consecutive rotations of the x-ray source. Therefore, $CTDI_w$ represents the average radiation dose over the x and y direction, whereas $CTDI_{vol}$ represents the average radiation dose over the x, y and z directions (Goldman, 2007).

Dose Length Product (DLP) is the measure of ionizing radiation exposure during the entire acquisition of images (Habib-Geryes, 2019).

$$DLP \text{ (mGy.cm)} = CTDI_{vol} \times \text{irradiated length (slice thickness} \times \text{number of slices)} \quad (3)$$

in centimeters.

$CTDI_w$ and $CTDI_{vol}$ are independent of scan length for determining the total energy absorbed whereas DLP is proportional to scan length.

The CT machine input parameters are kilovoltage peak (kV) and milliampere-second (mAs). Kilovoltage peak (kV) is the peak voltage applied to the x-ray tube. It determines the highest energy of the x-ray photon. It is responsible for the acceleration of electrons from the cathode to the anode.

In Computed Tomography, reducing kV can be an effective means of reducing the radiation dose imparted during an examination. As a rule of thumb, the radiation dose changes with the square of kV, thus a reduction in kV from 120 to 100 reduces radiation dose by 33% while a further reduction to 80kV can reduce dose by 65%. Nevertheless, unlike reductions in mAs, which have a linear and relatively predictable effect on image noise and contrast-to-noise ratios, decreases in kV can result to non-linear, exponential increases in image noise (Feuchtner et al., 2010).

Tube Loading (mAs) is a measure of radiation produced over a period of time in seconds via an x-ray tube. An increase in current (mA) results in a higher production of electrons in the x-ray tube. This will therefore increase the quantity of radiation produced. Increase in mAs improves image quality and decreases image noise. The relationship between the tube loading and patient dose is essentially linear. Increase in mAs results in a comparable increase in patient dose (Bushong, 2013).

3. MATERIAL AND METHODS

A retrospective study involving collection of various exposure parameters of Cranial and Abdomino - pelvic CT scan post investigation with a Toshiba Aquilion 64 slice CT machine at the Radiology Department, University College Hospital, Ibadan. The study population comprised of patients with medical request for cranial or abdomino-pelvic CT that report for the investigation within the period of the study at the CT unit of the Radiology Department,

UCH, Ibadan. The assessment involved adult Patients with medical request, who were scanned for cranial or Abdominal – pelvic CT, and exclude; children, unconscious patients, very ill patients, CT patients with high risk of contrast media administration and patients with other CT examination request.

The sample size was calculated considering the population of all patients whose CT (cranial and abdomino-pelvic) procedures were within the interval of April – June, 2019. The sample size was determined using Taro Yamane formula (Rafael, 2014). The formula is stated in equation 4.

$$n = N(1 + Ne^2) \quad (4)$$

Where N signifies the minimum sample size, e is the significant level (0.05%), resulted to 73 sample size.

$$n = 73(1 + 73 (0.05)^2) = 61.7 \quad (5)$$

Exposure parameters such as kV, mAs, Scan length, and Slice thickness were obtained from the CT scanner. The resultant dose descriptors: Volume Computed Tomography Dose Index $CTDI_{vol}$ and Dose Length Product (DLP) were obtained from the CT machine patients' data base. Total mAs and total scan time including patient's demographic data: age, sex and weight were also recorded. A total of 61 patients' data that underwent CT examination were retrospectively studied. Thirty-nine (39) patients underwent cranial examination and twenty-two (22) patients underwent Abdomino-pelvis examination. In the cranial category, there were 21 male and 18 female patients and in the Abdomino- pelvic category there were 8 male and 14 female patients. The mean age of patients for cranial examination was 52 ± 12 years while that of Abdomino-pelvic examination was 57 ± 3 years.

The data collected was analyzed descriptively (mean, range and standard deviation) using Statistical Package for Social Science (SPSS-27). Each of the scan parameters (kv, mAs, weight, scan length) were correlated with the dose descriptors ($CTDI_{vol}$ and DLP) using Pearson Product – Moment Correlation.

The exposure parameters; kV, mAs, slice thickness, for both procedures are presented in table 1. The acquisition detector configuration is also stated in the table, which is the product of the number of detector rows (N) and the slice thickness (T). Pitch is the ratio of the table movement per rotation and detector configuration.

In this study the pitch was 21 for cranial and 53 for abdomino- pelvis CT examinations.

4. RESULTS AND DISCUSSION

Table1: Exposure Parameters for Cranial CT and Abdomino-pelvis CT Examinations

| Exposure Parameters | Cranial CT | Abdomino-pelvis CT |
|------------------------------------|-------------------|---------------------------|
| Range of Tube Loading (mAs) | 200– 250 | 263-381 |
| Range of Tube Voltage (kV) | 100-120 | 100-120 |
| Acquisition Detector Configuration | 32 x 0.5mm | 32 x 1.0mm |
| Slice Thickness | 0.5mm | 1 mm |
| Rotation Time | 0.4s | 0.4s |
| Pitch | 21 | 53 |
| Range of Patients' weight (kg) | 42.9-98.8 | 46.7-93.1 |
| Range of Patients' Age (yrs) | 18–89 | 18–77 |
| Scan Length | 19-26 | 36.0 -46.5 |
| Total mAs | 8475-15300 | 3395-15285 |
| Total $CTDI_{vol}$ | 135.6-215.5 | 26.1-124.4 |
| Total DLP | 2631.0 - 6542.8 | 1147.5 - 5441.8 |

A comparison between kV, mAs, Scan length, weight and the dose descriptors CTDI_{vol} and DLP was done. Correlation was performed using Pearson's Correlation to establish the degree of linear relationship between two variables which is the Scan input parameters with the CTDI_{vol} and DLP

The relationships between weight and Total CTDI_{vol} with weight and DLP are represented with a scatter plot shown in Figure 1a and 1b respectively. There was a very weak correlation between weight and Total CTDI_{vol} ($r = .117, n = 39, p = > .05$). Similarly, there was no significant correlation between weight and DLP ($r = .285, n = 39, p = > .05$).




Fig. 1a: Scatter Plot of Weight and Total CTDI_{vol} in Cranial CT




Fig.1b: Scatter Plot of Weight and DLP in Cranial CT

The relationships between kV and Total CTDI_{vol} with kV and DLP are shown in Figure 2a and 2b. There was a significant correlation between the variables. There was a moderate positive correlation between kV and Total CTDI_{vol} ($r = .625, n = 39, p < .01$), which indicated a linear relationship, while the relationship between kV and DLP showed a strong positive correlation ($r = .767, n=39, p < .01$).




Fig. 2a: Scatter Plot of kV and Total CTDI_{vol} in Cranial CT




Fig. 2b: Scatter Plot of kV and DLP in cranial CT

The relationships between mAs and Total CTDI_{vol} with mAs and DLP are shown in Figure 3a and 3b. There was a significant correlation between the variables, with an r-value of 0.555 and .843 for CTDI_{vol} and DLP respectively. Figure 3a shows a moderate positive correlation between mAs and Total CTDI_{vol}, ($r = .555$, $n = 39$, $p < .01$), which indicated a linear relationship. Figure 3b shows a strong positive correlation between mAs and DLP ($r = .843$, $n = 39$, $p < .01$).




Fig. 3a: Scatter plot of mAs and Total CTDI in Cranial CT




Fig. 3b: Scatter plot of mAs and DLP in Cranial CT

The Weight and Total CTDI_{vol} was correlated for Abdomino-Pelvic Scan and there was a strong positive correlation between the two variables significantly ($r = .701$, $n = 22$, $p < .01$), which indicated a linear relationship. This is shown in figure 4a. The relationship between weight and DLP also had a strong positive correlation, significantly ($r = .817$, $n = 22$, $p < .01$). Figure 4b shows a scatter plot correlating weight and DLP.




Fig. 4a: Scatter Plot of Weight and Total CTDI_{vol} in Abdomino-Pelvic CT




Fig. 4b: Scatter Plot of Weight and DLP in Abdomino-Pelvic CT

The relationships between kV and Total CTDI_{vol} with kV and DLP are shown in Figure 5a and 5b. The kV and Total CTDI_{vol} was correlated and a strong positive correlation was observed between the two variables, significantly ($r = .821$, $n = 22$, $p < .01$). The kV and DLP also had a strong positive correlation, significantly ($r = .839$, $n=22$, $p < .01$).




Fig. 5a: Scatter Plot of kV and Total CTDI_{vol} in Abdomino- pelvic CT




Fig. 5b: Scatter Plot of kV and DLP in Abdomino-pelvic CT

The relationships between mAs and Total CTDI_{vol} with mAs and DLP are shown in Figure 6a and 6b. There was a significant correlation between the variables, with an r-value of .837 and .839 for CTDI_{vol} and DLP respectively, which indicated a linear relationship.




Fig. 6a: Scatter Plot of mAs and Total CTDI_{vo} in Abdomino- pelvic CT examination.




Fig. 6b: Scatter Plot of mAs and DLP in in Abdomino- pelvic CT examination

5. CONCLUSION

The results and findings obtained from this study are in line with the findings in the literatures. The dose metrics (CTDI_{vol} and DLP) for both Cranial and Abdomino-Pelvic CT scans had linear relationships with kV, mAs and scan length. Increase in these parameters however increases the dose metrics and patient dose. In order to reduce patient dose, the scan acquisition parameters (kV, mAs and scan length) need to be reduced. This finding confirms the need to optimize the CT scan acquisition parameters, which will be useful for CT scan radiographers, as part of dose reduction techniques. It is expected that a reduction of the scan input parameters will influence the image quality. Therefore, the effect of reduction of these parameters on image quality should be investigated in future studies.

ETHICAL APPROVAL

Ethical approval was granted by the University College Hospital, Ibadan, before commencing the research work.

REFERENCES

1. Anam C, Haryanto F, Widita R, Arif I, Fujibuchi T and Dougherty G. A size-specific effective dose for patients undergoing CT examinations. *J. Phys. Conf. Ser.* 2019; DOI 10.1088/1742-6596/1204/1/012002.
2. Aweda MA and Arogundade RA. Patient Dose Reduction Methods in Computerised Tomography Procedures: A review. *International Journal of Physical Sciences.* 2007; 2(1): 001-009.
3. Bjorkdahl P and Nyma U. Using 100 – instead of 120 kVp Computed Tomography to diagnose pulmonary embolism almost halves the radiation dose with preserved diagnostic quality. *Acta Radiologica.* 2010; 51(3): 260-270.
4. Bushong SC. Radiologic science for technologists St. Louis (3rded) Mosby. 2013: Elsevier.
5. Christner JA, Braun NN, Jacobsen MC, Carter RE, Kofler JM and McCollough CH. Size-Specific Dose Estimates for adult patients at Computed Tomography of the Torso. *Radiology.* 2012; 265 (3): 841-847.
6. Feuchtner GM, Jodocy D and Klauser A. Radiation Dose Reduction by using 100 –KV tube voltage in cardiac 64 – slice computed tomography: a comparative study. *Journal of Radiology.* 2010; (75): 51-56.
7. Goldman LW. Principles of Computed Tomography: Radiation Dose and Image Quality. *Journal of Nuclear Medicine Technology.* 2007; 35, (4): 213-225.
8. Habib-Geryes B, Hornbeck A, Jarrige V, Pierrat N, Ducou Le Pointe H, et al.. Patient dose evaluation in computed tomography: A French national study based on clinical indications. *Physica Medica,* 2019; (61):18-27.
9. Huda W. Radiation Dosimetry in CT: the role of the manufacturer. *Imaging Medical.* 2011; 3(2): 241-259.
10. Israel GM, Cicchiello L, Brink J and Huda W. Patient size and radiation exposure in thoracic, pelvic and abdominal computed tomography examinations performed with automatic exposure control. *AJR Am J Roentgenol.* 2010; 195 (6): 1342-1346.

11. Rafael C. How to Calculate Sample Size Using Taro Yamane's Formula. 2014; <https://www.nairaproject.com/blog/its-taro-yamane-not-yaro.htm>.
12. Smith-Bindman R, Miglioretti DL, Larson EB. Rising use of diagnostic medical imaging in a large integrated health system. *Health Aff Millwood* 2008; 27(6):1491-1502.
13. Yu L, Liu X, Leng S, Kofler JM, Ramirez-Giraldo JC., Qu M., Christner J., et al. Radiation dose reduction in Computed Tomography: Techniques and Future Perspective. *Imaging Medical*. 2009; 1(1): 65-84.
14. Zarb F, Ranford L, and Mc Entee MF. Antero-posterior diameter shows the strongest correlation with CTDI and DLP in abdominal and chest Computed Tomography. *Radiation Protection Dosimetry*. 2010; (2), 1-8.

UNDER PEER REVIEW