

# INCUBATORS in NEONATAL MEDICINE

## Abstract

The incubator for children to improve the survival chances of premature and immature newborns was developed in France as early as 1857. The first device in the United States was built by William Champion Deming at the State Emigrant Hospital on Ward's Island, New York. The first baby placed in it was Edith Eleanor McLean, who weighed 1106 grams at birth on September 7, 1888. The device was heated by 57 liters of water. Precursors to these devices, which mimic the consistent temperature in the womb, were the Ruehl cradle in Moscow in 1835 and the "warm bath" introduced by Credé in Leipzig in 1864. The technology of that time is incomparable to the technology of today's incubators and microbiological incubators.

**Keywords:** immature newborns, Transport incubators, microbiological incubators, gestation

## Introduction

Roughly every 11th child in Germany is a premature baby, babies born before reaching the 37th week of pregnancy. Many of these children spend their first weeks or months of life in the hospital in an incubator. The first incubators were developed in France in the mid-19th century. Since that time, when a lot of warm water flowed through the incubator to warm the bed, the technological advancements have significantly improved. However, then and now, an incubator serves to protect the newborn from germs and creates an environment similar to the mother's womb, where the tiniest babies can survive safely. At first glance, the infant incubator may seem very impersonal and cold to you: the incubator is a transparent box made of plexiglass with or without a cover. The box has openings on the sides through which the baby can be cared for. The incubator provides the tiniest babies with a protective environment. The baby and its vital functions can be monitored at all times. An incubator, colloquially known as a hatchery, and in medical terminology also called a *couveuse* (French for hatchery), is a medical device that creates and maintains controlled external conditions for various incubation and growth processes (3,8,17,22,67). Specifically, an incubator generates a microclimate with tightly regulated air humidity and temperature. The *couveuse* introduced in 1881 by Étienne Stéphane Tarnier (1828–1897) was an incubator for weak newborns (2,3). Transport incubators are used in inter-hospital transfers, which involve transporting premature and critically ill newborns from one hospital to a more suitable clinic with a specialized department for neonatal care. Transport incubators are designed to meet the specific requirements of incubator transport, ensuring mobility and quick and safe loading. The transport incubator must be capable of heating, enriching the air inside with oxygen, connecting a ventilation bag of appropriate size for the patient, and ensuring cleanliness and hygiene. The first transport incubators were introduced in the mid-1950s by Hermann Hilber in Munich (5,9,11,47). Intensive care transport incubators are used when the vital functions of premature or newborn infants are at risk. These incubators are equipped with advanced intensive care therapy options, such as suction devices, respirators (ventilators), infusion pumps, and additional monitoring equipment for patient data (5,9,12,24).

## Incubation process

Incubators create controlled conditions, a closed climate room, in which specific growth processes can occur (1-75). In addition to the incubators with constant high temperature and humidity used for premature infants, incubators are also referred to as growth chambers in medical laboratories. These microbiological incubators are used for culturing disease-causing organisms or storing blood and human tissue. In this context, the incubator for microbiology itself plays a significant role (23,45,66). In the neonatal unit, air-conditioned and mobile incubators are used not only for the care of premature and critically ill newborns but also for their transport, and in this form, they are also referred to as transport incubators. Every type of medical incubator creates controlled and optimizable external conditions by adjusting factors such as temperature within the devices to a specific growth or incubation process. With the advancement of technology, incubators have also evolved, and now highly specialized and advanced incubators are available for various applications. An incubator is a closed climate room, ensuring the survival of premature infants. The function of medical incubators is to generate ideal growth conditions and support a growth process. Infant incubators are the most well-known type of incubators. In the context of newborns, transport incubators are also frequently used, allowing for the inter-hospital transfer of premature or critically ill babies. 20 million preterm babies are born worldwide each year, whereas 4 million die in the first months worldwide. 63000 children are born as preterm infants in Germany. 8000 children are born before 30 weeks of gestation and weighing less than 1,500 grams in Germany. The viability limit is 22-24 weeks of pregnancy. Treatment will be performed only in specialized perinatal germancenters. Therapy costs for

"extreme low birthweight preterm" are 90000 euros per year. In 2021 an incubator that operates without electricity, by Fabian Nest, for Sub-Saharan Africa. Different types of incubators are available, a neonatal incubator, an Open-Box incubator, a Close Box incubator and a double-walled incubator (1-75). North America dominates the market. 1:10 of newborns are preterm births. This high number of preterms demand for increasing incubator necessity. In general, a transport may be necessary when the baby needs to be transferred from the birthplace to a more specialized hospital that can better meet its needs. Typically, the transport in this scenario is carried out as part of a baby ambulance operation (34,56,66). The newborn is placed in the baby ambulance within the transport incubator, which allows for both mobility and a safe, hygienic, and quick loading of the baby. Like all infant incubators, transport incubators are heated and enriched with oxygen. Additionally, a manual resuscitator may be connected to the incubator to provide continuous oxygen to critically ill newborns. In addition to conventional transport incubators, intensive care transport incubators are also available, especially for newborns with compromised vital functions. These specialized boxes are equipped with intensive care connections for a suction device or an infusion pump. Monitors for monitoring body functions can also be connected to these types of incubators. Incubator therapy for newborns carries some risks. Babies rely on a relatively constant body temperature, which may not be fully maintained by the incubator treatment. Therefore, the infant's body temperature must be regularly monitored, ideally rectally (20,45,56,57). Hypothermia, or cooling of the newborn, can occur. On the other hand, hyperthermia, or overheating of the baby, can also occur: as a result of overheating, the baby loses a lot of fluid, potentially leading to electrolyte imbalances. Hyperventilation or tachycardia are also possible. In general, the correct temperature can be set and adjusted on modern devices through skin temperature regulation. However, for extremely premature infants with high fluid loss due to shock or infections, it may be difficult to accurately measure the skin temperature or temperature changes of the young patient. Additionally, oxygen supply in an incubator can pose risks for a newborn. Inadequate oxygen supply due to faulty measuring devices can lead to apnea and ultimately brain damage. In premature infants, overdosing of oxygen can also occur due to measurement errors, potentially damaging the newborn's eyes. Since oxygen increases the risk of fire, personnel must also ensure the elimination of fire hazards in the immediate vicinity of the incubator (31,55,67).

### **Stability of the incubator**

Stability of the incubator plays an important role in interhospital transfer of the babies, the transport from station to station (34,44). A battery provides this mobility status. A CO-Sensor, carbon monoxide sensor, measures regularly the CO-concentration in the incubator. Temperature constancy is very important for the baby. Integrated scales for weight measurement and constant oxygen supply with a sufficient oxygen saturation percentage is of utmost necessity. There is an infection risk in warm and humid conditions (23,66,67,70). Humidification control is also very important. Transepidermal water loss inversely correlates with gestational age at birth (TEWL) (4,7,9,13). Recent studies have focused on noise reduction in the incubator ("neurodevelopmental dungeon"). There have been noises from the oxygen airflow, from the micro-humidification systems, from a loud environment. Permanent hearing impairments in preterm infants are recently described. Moreover a delayed language acquisition was found. Therefore stability of a closed climate chamber is crucial. The incubator has a resonance frequency of 100 Hz. External noises sound tonal, booming, and muffled inside the incubator, initiating reflex status and sleep behavior changes. Due to the oxygen amount, a fire risk is known in contact with flammable materials on the incubator. Factors influencing the market of incubator production include the prevalence of preterm births, low birth weight, a growing research community, Covid-19 in pregnant women, maternal age and a market growth until 2028. High costs with a missing coverage in developing countries is challenging. The possibility of X-ray application in the incubator is appreciated in different newborn diseases. Artificial Intelligence with a Fuzzy controller (FUZZY LOGIC CONTROL) plays a more and more important role with monitoring of actual and target values, corrections to desired values immediately and moreover an incubator parameter control for controlling multiple input and output variables with highly variable parameters (34,74,75). Phototherapy with LED lights/sensors and light-emitting diodes and light-dependent resistors are obligatory features of an incubator of younger age. Heating enrolls 36-38 degrees, whereas the heater converts electrical energy into thermal energy (3,7,8,22).

### **Conclusion**

The future will add new technology like artificial intelligence (AI) into the incubator system to control all parameters and all changes in parameters timely as close as possible (76-79). A fully operational incubator with precise control over temperature, humidity, and airflow was designed and evaluated in a project (79). Alongside the incubator development, a heuristic simulation was developed to fine-tune and test the Mamdani fuzzy logic controller (79). The controller was subsequently implemented in the incubator prototype (79). Premature birth is a global issue, with thermoregulation posing a significant challenge for premature infants who are often placed in incubators for convective heating (81). Current methods of control involve sensing either the incubator air temperature or the infant's skin temperature for heat regulation (81). However, using only one of these parameters can result in fluctuations in either the air or skin temperature. A further study aimed to investigate the feasibility of incorporating both the incubator air temperature and the infant's skin temperature in a fuzzy logic control system for heating regulation (81).

The control system was tested using a mathematical model of the infant-incubator system, confirming that traditional control methods can lead to temperature fluctuations or slow rise times in core temperature (81). The fuzzy logic control demonstrated a smoother and more effective regulation with the desired rise time (81,82). Another papers present the application of adaptive fuzzy sliding mode control (AFSMC) for a respiratory system to assist the patients facing difficulty in breathing (83). Compared to neonatal care, this adaptive sliding mode control (AFSMC) could play an important role in controlling parameters of newborn in an incubator (83,84).

## References

- (1) Featherstone PJ, Ball CM. Neonatal incubators. *Anaesth Intensive Care*. 2023 Mar;51(2):84-87. doi: 10.1177/0310057X221148277. PMID: 36872907.
- (2) Denk D, Greten FR. Inflammation: the incubator of the tumor microenvironment. *Trends Cancer*. 2022 Nov;8(11):901-914. doi: 10.1016/j.trecan.2022.07.002. Epub 2022 Jul 28. PMID: 35907753.
- (3) Chavignon M, Reboux M, Tasse J, Tristan A, Claris O, Laurent F, Butin M. Persistent microbial contamination of incubators despite disinfection. *Pediatr Res*. 2021 Dec;90(6):1215-1220. doi: 10.1038/s41390-021-01407-8. Epub 2021 Feb 24. PMID: 33627818.
- (4) LeBlanc MH. Thermoregulation: incubators, radiant warmers, artificial skins, and body hoods. *Clin Perinatol*. 1991 Sep;18(3):403-22. PMID: 1934849.
- (5) Wright WE, Shay JW. Inexpensive low-oxygen incubators. *Nat Protoc*. 2006;1(4):2088-90. doi: 10.1038/nprot.2006.374. PMID: 17487199.
- (6) Reboux M, Chavignon M, Tristan A, Plaisant F, Laurent F, Butin M. Disinfection of incubators in neonatal intensive care units: impact of steam pulverization on bacterial colonization. *Antimicrob Resist Infect Control*. 2023 Mar 16;12(1):18. doi: 10.1186/s13756-023-01226-y. PMID: 36927466; PMCID: PMC10022080.
- (7) Harding GH. Double-wall infant incubators. *Crit Care Med*. 1984 Mar;12(3):205. doi: 10.1097/00003246-198403000-00013. PMID: 6697746.
- (8) Laroia N, Phelps DL, Roy J. Double wall versus single wall incubator for reducing heat loss in very low birth weight infants in incubators. *Cochrane Database Syst Rev*. 2007 Apr 18;2007(2):CD004215. doi: 10.1002/14651858.CD004215.pub2. PMID: 17443543; PMCID: PMC8767785.
- (9) Yumoto K, Iwata K, Sugishima M, Yamauchi J, Nakaoka M, Tsuneto M, Shimura T, Flaherty S, Mio Y. Unstable osmolality of microdrops cultured in non-humidified incubators. *J Assist Reprod Genet*. 2019 Aug;36(8):1571-1577. doi: 10.1007/s10815-019-01515-9. Epub 2019 Jul 2. PMID: 31267335; PMCID: PMC6707990.
- (10) Tugwell-Allsup J, Kenworthy D, England A. Mobile chest imaging of neonates in incubators: Optimising DR and CR acquisitions. *Radiography (Lond)*. 2021 Feb;27(1):75-80. doi: 10.1016/j.radi.2020.06.005. Epub 2020 Jul 4. PMID: 32636056.
- (11) Fukuyama T, Arimitsu T. Use of access port covers in transport incubators to improve thermoregulation during neonatal transport. *Sci Rep*. 2023 Feb 23;13(1):3132. doi: 10.1038/s41598-023-30142-9. PMID: 36823206; PMCID: PMC9950442.
- (12) Bergez L, Jourdain G, De Luca D. Noise Produced by Neonatal Ventilators Inside and Outside of the Incubators. *Respir Care*. 2023 Nov 25;68(12):1693-1700. doi: 10.4187/respcare.10989. PMID: 37147103; PMCID: PMC10676250.
- (13) Marik PE, Fuller C, Levitov A, Moll E. Neonatal incubators: a toxic sound environment for the preterm infant?\*. *Pediatr Crit Care Med*. 2012 Nov;13(6):685-9. doi: 10.1097/PCC.0b013e31824ea2b7. PMID: 22791088.
- (14) Boyd H, Brand MC, Hagan J. Care of 500-1500 Gram Premature Infants in Hybrid Incubators. *Adv Neonatal Care*. 2017 Oct;17(5):381-389. doi: 10.1097/ANC.0000000000000422. PMID: 28799947.
- (15) Rossetti V, Abriel H. The NCCR TransCure: An Incubator for Interdisciplinary Research. *Chimia (Aarau)*. 2022 Dec 21;76(12):992-995. doi: 10.2533/chimia.2022.992. PMID: 38069793.
- (16) Yu X, Gujjula S, Kuo SM. Active noise control for infant incubators. *Annu Int Conf IEEE Eng Med Biol Soc*. 2009;2009:2531-4. doi: 10.1109/IEMBS.2009.5334818. PMID: 19964974.
- (17) Delanaud S, Decima P, Pelletier A, Libert JP, Durand E, Stephan-Blanchard E, Bach V, Tourneux P. Thermal management in closed incubators: New software for assessing the impact of humidity on the optimal incubator air temperature. *Med Eng Phys*. 2017 Aug;46:89-95. doi: 10.1016/j.medengphy.2017.06.002. Epub 2017 Jun 21. PMID: 28645849.
- (18) Mortazavi SMJ. Re: Are electromagnetic fields in incubators a risk factor for autism? *Acta Paediatr*. 2017 Dec;106(12):2063. doi: 10.1111/apa.14055. Epub 2017 Sep 24. PMID: 28862780.
- (19) Gavinelli R, Sacerdote CB. Il rumore nelle incubatrici [Noise in incubators]. *Minerva Pediatr*. 1978 Jan 31;30(2):185-9. Italian. PMID: 565460.
- (20) Colareta Ugarte U, Prazad P, Puppala BL, Schweig L, Donovan R, Cortes DR, Gulati A. Emission of volatile organic compounds from medical equipment inside neonatal incubators. *J Perinatol*. 2014 Aug;34(8):624-8. doi: 10.1038/jp.2014.65. Epub 2014 Apr 24. PMID: 24762411.

- (21) de Carvalho M, Torrao CT, Moreira ME. Mist and water condensation inside incubators reduce the efficacy of phototherapy. *Arch Dis Child Fetal Neonatal Ed.* 2011 Mar;96(2):F138-40. doi: 10.1136/adc.2010.189423. Epub 2010 Sep 16. PMID: 20847196.
- (22) Delanaud S, Decima P, Pelletier A, Libert JP, Stephan-Blanchard E, Bach V, Tourneux P. Additional double-wall roof in single-wall, closed, convective incubators: Impact on body heat loss from premature infants and optimal adjustment of the incubator air temperature. *Med Eng Phys.* 2016 Sep;38(9):922-8. doi: 10.1016/j.medengphy.2016.05.010. Epub 2016 Jul 4. PMID: 27387899.
- (23) Robertson A, Cooper-Peel C, Vos P. Sound transmission into incubators in the neonatal intensive care unit. *J Perinatol.* 1999 Oct-Nov;19(7):494-7. doi: 10.1038/sj.jp.7200258. PMID: 10685297.
- (24) Fackler JC. Incubator noise: don't throw out the baby with the bathwater\*. *Pediatr Crit Care Med.* 2012 Nov;13(6):699-700. doi: 10.1097/PCC.0b013e318253c828. PMID: 23128594.
- (25) Bell EF, Rios GR. Air versus skin temperature servocontrol of infant incubators. *J Pediatr.* 1983 Dec;103(6):954-9. doi: 10.1016/s0022-3476(83)80729-x. PMID: 6644435.
- (26) Yu H, Yin J, Zhang L, Wang W. [Study on Adverse Events of Infant Incubator Products]. *Zhongguo Yi Liao Qi Xie Za Zhi.* 2021 Jun 8;45(3):335-339. Chinese. doi: 10.3969/j.issn.1671-7104.2021.03.022. PMID: 34096248.
- (27) Wheldon AE, Rutter N. The heat balance of small babies nursed in incubators and under radiant warmers. *Early Hum Dev.* 1982 Apr;6(2):131-43. doi: 10.1016/0378-3782(82)90100-1. PMID: 7094851.
- (28) Bess FH, Peek BF, Chapman JJ. Further observations on noise levels in infant incubators. *Pediatrics.* 1979 Jan;63(1):100-6. PMID: 440786.
- (29) Deguines C, Dégrugilliers L, Ghyselen L, Chardon K, Bach V, Tourneux P. Impact of nursing care on temperature environment in preterm newborns nursed in closed convective incubators. *Acta Paediatr.* 2013 Mar;102(3):e96-e101. doi: 10.1111/apa.12109. Epub 2013 Jan 4. PMID: 23190392.
- (30) Robertson A, Stuart A, Walker L. Transmission loss of sound into incubators: implications for voice perception by infants. *J Perinatol.* 2001 Jun;21(4):236-41. doi: 10.1038/sj.jp.7210531. PMID: 11533840.
- (31) Kurimoto T, Ibara S, Ishihara C, Naito Y, Hirakawa E, Yamamoto T. Incubator humidity and temperature control in infants born at 22-23 weeks' gestation. *Early Hum Dev.* 2022 Mar;166:105550. doi: 10.1016/j.earlhumdev.2022.105550. Epub 2022 Jan 31. PMID: 35151106.
- (32) French NA. Modeling incubation temperature: the effects of incubator design, embryonic development, and egg size. *Poult Sci.* 1997 Jan;76(1):124-33. doi: 10.1093/ps/76.1.124. PMID: 9037699.
- (33) Rodarte MDO, Fujinaga CI, Leite AM, Salla CM, Silva CGD, Scochi CGS. Exposure and reactivity of the preterm infant to noise in the incubator. *Codas.* 2019 Nov 7;31(5):e20170233. Portuguese, English. doi: 10.1590/2317-1782/20192017233. PMID: 31721912.
- (34) Riera J, Jiménez Mñoz JM. Noticiaretrospectiva de las primeras incubadoras pediátricas [Incubators in XIX century (author's transl)]. *An Esp Pediatr.* 1974 Nov-Dec;7(6):590-3. Spanish. PMID: 4617538.
- (35) Bekiesińska-Figatowska M, Rutkowska M, Stankiewicz J, Krupa K, Iwanowska B, Romaniuk-Doroszeńska A, Szkudlińska-Pawlak S, Duczowska A, Duczowski M, Bragoszeńska H, Mądzik J, Kwaśniewicz P, Cabaj A, Helwich E. Neonatal brain and body imaging in the MR-compatible incubator. *Adv Clin Exp Med.* 2019 Jul;28(7):945-954. doi: 10.17219/acem/94155. PMID: 31111693.
- (36) Bearer CF. Electromagnetic fields and infant incubators. *Arch Environ Health.* 1994 Sep-Oct;49(5):352-4. doi: 10.1080/00039896.1994.9954986. PMID: 7944566.
- (37) Aynsley-Green A, Robertson NR, Rolfe P. Air temperature recordings in infant incubators. *Arch Dis Child.* 1975 Mar;50(3):215-9. doi: 10.1136/adc.50.3.215. PMID: 1147654; PMCID: PMC1544499.
- (38) Tiam Kapen P, Youssoufa M, Foutse M, Dongmeza Koudjou J, Mkankam Kamga FP. A multi-function neonatal incubator for low-income countries: Implementation and ab initio social impact. *Med Eng Phys.* 2020 Mar;77:114-117. doi: 10.1016/j.medengphy.2019.10.021. Epub 2020 Jan 11. PMID: 31937436.
- (39) Kingma E, Finn S. Neonatal incubator or artificial womb? Distinguishing ectogestation and ectogenesis using the metaphysics of pregnancy. *Bioethics.* 2020 May;34(4):354-363. doi: 10.1111/bioe.12717. Epub 2020 Apr 5. PMID: 32249443.
- (40) Durand O. Equipment survey: infant respirators and respirator-incubators. *Respir Ther.* 1979 Jan-Feb;9(1):46. PMID: 10297351.
- (41) Lagercrantz H. Are extremely preterm born children with autism the victims of too much isolation in the incubator? *Acta Paediatr.* 2017 Aug;106(8):1246-1247. doi: 10.1111/apa.13874. Epub 2017 May 15. PMID: 28414876.
- (42) Haffner ME. Malfunction of neonatal incubator. *JAMA.* 1982 May 7;247(17):2372. PMID: 7069893.
- (43) Zaylaa AJ, Rashid M, Shaib M, El Majzoub I. A Handy Preterm Infant Incubator for Providing Intensive Care: Simulation, 3D Printed Prototype, and Evaluation. *J Healthc Eng.* 2018 May 10;2018:8937985. doi: 10.1155/2018/8937985. PMID: 29861884; PMCID: PMC5971329.
- (44) Tugwell-Allsup J, England A. A systematic review of incubator-based neonatal radiography - What does the evidence say? *Radiography (Lond).* 2020 May;26(2):167-173. doi: 10.1016/j.radi.2019.09.009. Epub 2019 Oct 16. PMID: 32052784.

- (45) Templeman MC, Bell EF. Head insulation for premature infants in servocontrolled incubators and radiant warmers. *Am J Dis Child*. 1986 Sep;140(9):940-2. doi: 10.1001/archpedi.1986.02140230110046. PMID: 3740003.
- (46) Ren Q, Shao Y, Yu W. Pre-warming the double-lumen endobronchial tubes to facilitate intubation in incubator. *Br J Anaesth*. 2017 Jan;118(1):140-141. doi: 10.1093/bja/aew418. PMID: 28039253.
- (47) PICKEN JC Jr, BAURIEDEL WR. A low-temperature incubator. *Science*. 1950 Oct 6;112(2910):409. doi: 10.1126/science.112.2910.409. PMID: 14781781.
- (48) Blennow G, Svenningsen NW, Almquist B. Letter: Incubator noise. *Pediatrics*. 1975 Oct;56(4):617. PMID: 1165972.
- (49) Uithoven-Dolsma J, Cats BP, Tanke M, Hermanns J. Couveusekind: wat overkomt me nu? [Incubator child: what's happening to me now?]. *TijdschrZiekenverpl*. 1983 Apr 26;36(9):272-4. Dutch. PMID: 6552789.
- (50) Waffarn F, Hodgman JE. Mercury vapor contamination of infant incubators: a potential hazard. *Pediatrics*. 1979 Nov;64(5):640-2. PMID: 492838.
- (51) FadhillahNugraha P, Putra N, Ariantara B, Amin M. The use of beeswax as heating element in non-electric infant incubator. *J Med Eng Technol*. 2017 Nov;41(8):593-599. doi: 10.1080/03091902.2017.1382586. Epub 2017 Oct 18. PMID: 29043873.
- (52) Frankenberger RT, Bussmann O, Nahm W, Konecny E, Gortner L. MessungseitlicherHauttemperaturprofile von Frühgeborenen in InkubatorenmittelsThermographie [Measuring lateral skin temperature profile of premature infants in incubators with thermography]. *Biomed Tech (Berl)*. 1998 Jun;43(6):174-8. German. doi: 10.1515/bmte.1998.43.6.174. PMID: 9677758.
- (53) Gädeke R, Petersen P, de Liddle IW. Untersuchungen über die Geräuschbelastung von Säuglingen in Inkubatoren [Studies on noise stress caused by infant incubators (author's transl)]. *Monatsschr Kinderheilkd* (1902). 1979 Mar;127(3):144-8. German. PMID: 423903.
- (54) Erhani R, Dégrugilliers L, Lahana A, Glusko-Charlet A, Haraux E, Durand E, Tourneux P. Failing to meet relative humidity targets for incubated neonates causes higher heat loss and metabolic costs in the first week of life. *Acta Paediatr*. 2018 Jul;107(7):1177-1183. doi: 10.1111/apa.14063. Epub 2017 Oct 5. PMID: 28880399.
- (55) Ducker DA, Lyon AJ, Ross Russell R, Bass CA, McIntosh N. Incubator temperature control: effects on the very low birthweight infant. *Arch Dis Child*. 1985 Oct;60(10):902-7. doi: 10.1136/ad.60.10.902. PMID: 4062342; PMCID: PMC1777487.
- (56) Marks KH, Lee CA, Bolan CD Jr, Maisels MJ. Oxygen consumption and temperature control of premature infants in a double-wall incubator. *Pediatrics*. 1981 Jul;68(1):93-8. PMID: 7243514.
- (57) Christensson K, Bhat GJ, Amadi BC, Eriksson B, Höjer B. Randomised study of skin-to-skin versus incubator care for rewarming low-risk hypothermic neonates. *Lancet*. 1998 Oct 3;352(9134):1115. doi: 10.1016/S0140-6736(98)00028-2. PMID: 9798589.
- (58) Greenspan JS, Cullen AB, Touch SM, Wolfson MR, Shaffer TH. Thermal stability and transition studies with a hybrid warming device for neonates. *J Perinatol*. 2001 Apr-May;21(3):167-73. doi: 10.1038/sj.jp.7200513. PMID: 11503103.
- (59) Sinclair L, Crisp J, Sinn J. Variability in incubator humidity practices in the management of preterm infants. *J Paediatr Child Health*. 2009 Sep;45(9):535-40. doi: 10.1111/j.1440-1754.2009.01555.x. PMID: 19761481.
- (60) Breen PP, Buskila Y. Brainincubator: an incubation system to extend brain slice lifespan for use in neurophysiology. *Annu Int Conf IEEE Eng Med Biol Soc*. 2014;2014:4864-7. doi: 10.1109/EMBC.2014.6944713. PMID: 25571081.
- (61) Hellström-Westas L, Inghammar M, Isaksson K, Rosén I, Stjernqvist K. Short-term effects of incubator covers on quiet sleep in stable premature infants. *Acta Paediatr*. 2001 Sep;90(9):1004-8. doi: 10.1080/080352501316978075. PMID: 11683187.
- (62) Liu HC. [Centralized management strategy of the infant incubators in NICUs]. *Zhongguo Yi Liao Qi Xie Za Zhi*. 2005 Jul;29(5):371-3. Chinese. PMID: 16419950.
- (63) Christensson K, Siles C, Cabrera T, Belaustequi A, de la Fuente P, Lagercrantz H, Puyol P, Winberg J. Lower body temperatures in infants delivered by caesarean section than in vaginally delivered infants. *Acta Paediatr*. 1993 Feb;82(2):128-31. doi: 10.1111/j.1651-2227.1993.tb12622.x. PMID: 8477157.
- (64) Kim YH, Kwon CH, Yoo SC. Experimental and numerical studies on convective heat transfer in a neonatal incubator. *Med Biol Eng Comput*. 2002 Jan;40(1):114-21. doi: 10.1007/BF02347704. PMID: 11954698.
- (65) Reddy NP, Mathur G, Hariharan SI. Toward a fuzzy logic control of the infant incubator. *Ann Biomed Eng*. 2009 Oct;37(10):2146-52. doi: 10.1007/s10439-009-9754-6. Epub 2009 Jul 17. PMID: 19609677.
- (66) Cone TE Jr. The first published report of an incubator for use in the care of the premature infant (1857). *Am J Dis Child*. 1981 Jul;135(7):658-60. doi: 10.1001/archpedi.1981.02130310062020. PMID: 7018217.

- (67) Johanson RB, Malla DS, Tuladhar C, Amatya M, Spencer SA, Rolfe P. A survey of technology and temperature control on a neonatal unit in Kathmandu, Nepal. *J Trop Pediatr.* 1993 Feb;39(1):4-10. doi: 10.1093/tropej/39.1.4. PMID: 8445687.
- (68) Burunkaya M, Yucel M. Measurement and Control of an Incubator Temperature by Using Conventional Methods and Fiber Bragg Grating (FBG) Based Temperature Sensors. *J Med Syst.* 2020 Aug 27;44(10):178. doi: 10.1007/s10916-020-01650-2. PMID: 32856101.
- (69) Baker JP. The incubator controversy: pediatricians and the origins of premature infant technology in the United States, 1890 to 1910. *Pediatrics.* 1991 May;87(5):654-62. PMID: 2020510.
- (70) Riggs RJ. A baby incubator designed for conditions in Ghana. *J Med Eng Technol.* 1977 Jan;1(1):33-8. doi: 10.3109/03091907709161586. PMID: 839522.
- (71) Unno N, Kuwabara Y, Okai T, Kido K, Nakayama H, Kikuchi A, Narumiya Y, Kozuma S, Taketani Y, Tamura M. Development of an artificial placenta: survival of isolated goat fetuses for three weeks with umbilical arteriovenous extracorporeal membrane oxygenation. *Artif Organs.* 1993 Dec;17(12):996-1003. doi: 10.1111/j.1525-1594.1993.tb03181.x. PMID: 8110074.
- (72) Ginalski MK, Nowak AJ, Wrobel LC. A combined study of heat and mass transfer in an infant incubator with an overhead screen. *Med Eng Phys.* 2007 Jun;29(5):531-41. doi: 10.1016/j.medengphy.2006.07.011. Epub 2006 Oct 9. PMID: 17030142.
- (73) Tona K, Everaert N, Willemsen H, Gbeassor M, Decuypere E, Buyse J. Effects of interaction of incubator CO<sub>2</sub> levels and mixing hatching eggs of different embryo growth trajectory on embryo physiological and hatching parameters. *Br Poult Sci.* 2013;54(4):545-51. doi: 10.1080/00071668.2013.807907. PMID: 23906223.
- (74) Bell EF, Rios GR. A double-walled incubator alters the partition of body heat loss of premature infants. *Pediatr Res.* 1983 Feb;17(2):135-40. doi: 10.1203/00006450-198302000-00011. PMID: 6402753.
- (75) Aasen SE, Johnsson A, Bratlid D, Christensen T. Fifty-Hertz magnetic field exposures of premature infants in a neonatal intensive care unit. *Biol Neonate.* 1996;70(5):249-64. doi: 10.1159/000244374. PMID: 8955911.
- (76) Zhang N, Wood O, Yang Z, Xie J. AI-Guided Computing Insights into a Thermostat Monitoring Neonatal Intensive Care Unit (NICU). *Sensors (Basel).* 2023 May 5;23(9):4492. doi: 10.3390/s23094492. PMID: 37177696; PMCID: PMC10181714.
- (77) Jaschke AC, Bos AF. Concept and considerations of a medical device: the active noise cancelling incubator. *Front Pediatr.* 2023 Jul 3;11:1187815. doi: 10.3389/fped.2023.1187815. PMID: 37465419; PMCID: PMC10350684.
- (78) Gleichauf J, Hennemann L, Fahlbusch FB, Hofmann O, Niebler C, Koelpin A. Sensor Fusion for the Robust Detection of Facial Regions of Neonates Using Neural Networks. *Sensors (Basel).* 2023 May 19;23(10):4910. doi: 10.3390/s23104910. PMID: 37430829; PMCID: PMC10223875.
- (79) Zimmer DB, Inks AAP, Clark N, Sendi C. Design, Control, and Simulation of a Neonatal Incubator. *Annu Int Conf IEEE Eng Med Biol Soc.* 2020 Jul;2020:6018-6023. doi: 10.1109/EMBC44109.2020.9175407. PMID: 33019343.
- (80) Weber R, Cabon S, Simon A, Poree F, Carrault G. Preterm Newborn Presence Detection in Incubator and Open Bed Using Deep Transfer Learning. *IEEE J Biomed Health Inform.* 2021 May;25(5):1419-1428. doi: 10.1109/JBHI.2021.3062617. Epub 2021 May 11. PMID: 33646962.
- (81) Reddy NP, Mathur G, Hariharan SI. Toward a fuzzy logic control of the infant incubator. *Ann Biomed Eng.* 2009 Oct;37(10):2146-52. doi: 10.1007/s10439-009-9754-6. Epub 2009 Jul 17. PMID: 19609677.
- (82) Lin SC, Luo CH, Yeh TF. Fuzzy oxygen control system for the indirect calorimeter of premature infants. *J Med Eng Technol.* 2001 Jul-Aug;25(4):149-55. doi: 10.1080/03091900010022229. PMID: 11601441.
- (83) Mehedi IM, Shah HSM, Al-Saggaf UM, Mansouri R, Bettayeb M. Adaptive Fuzzy Sliding Mode Control of a Pressure-Controlled Artificial Ventilator. *J Healthc Eng.* 2021 Jun 23;2021:1926711. doi: 10.1155/2021/1926711. PMID: 34257849; PMCID: PMC8249163.
- (84) Hu Y, Meng J, Li G, Zhao D, Feng G, Zuo G, Liu Y, Zhang J, Shi C. Fuzzy Adaptive Passive Control Strategy Design for Upper-Limb End-Effector Rehabilitation Robot. *Sensors (Basel).* 2023 Apr 17;23(8):4042. doi: 10.3390/s23084042. PMID: 37112385; PMCID: PMC10146308.